

Analysis of Dicamba and 2,4-D by the Triple Quad LCMS-8060

No. LCMS-095



■ Executive Summary

Formulation assay including LOD, LOQ and robustness tests of dicamba and 2,4-D spiked into a commercial brand herbicide.

■ Background

The purpose of this analysis was to demonstrate the sensitivity of the LCMS-8060 when analyzing dicamba and 2,4-D in a solution of 100x diluted glufosinate herbicide that contains a high amount of surfactants and other potential mass spec interferants.

■ Method

The LC/MS/MS method was quickly developed with readily available mobile phases. Dicamba in methanol 0.01 and 0.025 ppb had a S/N of 3/1 and 10/1, respectively. 2,4-D in methanol at 0.0001 and 0.001 ppb had a S/N of 13/1 and 90/1, respectively. See Figures 1-4 for chromatograms

MRM and source conditions were optimized using LabSolutions Connect software and flow injection analyses. Parameters optimized include:

- Collision energy (rough and fine-tuned)
- Q1 Pre-rod voltage
- Q3 Pre-rod voltage

- Interface voltage (probe voltage)
- Collision gas pressure
- Nebulizing gas flow
- Drying gas flow
- Heating gas flow
- Interface temperature
- Desolvation line temperature
- Heating block temperature

LabSolutions Connect software chose the best parameters based upon the intensity of the peak. Only about 40 μ L of standard was used in the optimizations. See Figures 5-6 for LabSolutions Connect results.

A calibration curve was prepared in 100x diluted herbicide from 0.1 – 10 ppb dicamba and 0.01 – 1 ppb 2,4-D. An LOD sample at 0.01 ppb dicamba and 0.001 2,4-D was also analyzed. A duplicate curve was analyzed at the beginning and near the end of the analysis along with five sets of QCs (triplicate injections) and 134 injections of the mid-QC over time to show ruggedness. See Figures 7-8 (dicamba) and Figures 9-10 (2,4-D) for the LOD at injection 13 and injection 222, respectively. See Figures 11-12 (dicamba) and Figures 13-14 (2,4-D) for the LOQ at injection 16 and injection 225, respectively.

There were 134 injections of the mid-level QC over time with a % RSD of 1.87 and 3.16 for dicamba and 2,4-D, respectively. See Table 1 for more % RSDs. See Table 2 for curve accuracy values and Figures 15-16 for a graphical representation of the curves and QCs. See Figures 17-20 for a graphical representation of variance in the areas of the MQC injections as well as the LOD injections.




Figure 1: Limit of Detection for Dicamba (0.01 ppb)




Figure 2: Dicamba, 0.025 ppb




Figure 3: 2,4-D, 0.0001 ppb





Figure 4: 2,4-D, 0.001 ppb


Compound Name	Purity	Processor m/z	m/z1	CE1	Message	Interface Voltage	CID Gas	Nebulizing Gas	Drying Gas	Heating Gas	Interface Temp.	DL Temp	Heut Block Temp	User Name
1 Dicamba	-	219	175	7	The boundary v...	-5	210	2.5	8	6	200	250	200	System Administrator



Data File Name	Interface Voltage	CID Gas	Nebulizing Gas	Drying Gas	Heating Gas	Interface Temp	DL Temp	Heut Block Temp	User Name	Date Acquired
1 C:\LabSolutions\Data\Dicamba 2-4 D\Optimization -connect.20171017_160752\Data.Dic...	-1	270	2	10	10	300	250	400	System Administrator	2017/10/17 16:09
2 C:\LabSolutions\Data\Dicamba 2-4 D\Optimization -connect.20171017_160752>Data.Dic...	-1.5	270	2	10	10	300	250	400	System Administrator	2017/10/17 16:10
3 C:\LabSolutions\Data\Dicamba 2-4 D\Optimization -connect.20171017_160752>Data.Dic...	-2	270	2	10	10	300	250	400	System Administrator	2017/10/17 16:12
4 C:\LabSolutions\Data\Dicamba 2-4 D\Optimization -connect.20171017_160752>Data.Dic...	-3	270	2	10	10	300	250	400	System Administrator	2017/10/17 16:14
5 C:\LabSolutions\Data\Dicamba 2-4 D\Optimization -connect.20171017_160752>Data.Dic...	-4	270	2	10	10	300	250	400	System Administrator	2017/10/17 16:15
6 C:\LabSolutions\Data\Dicamba 2-4 D\Optimization -connect.20171017_160752>Data.Dic...	-5	270	2	10	10	300	250	400	System Administrator	2017/10/17 16:17
7 C:\LabSolutions\Data\Dicamba 2-4 D\Optimization -connect.20171017_160752>Data.Dic...	-5	210	2	10	10	300	250	400	System Administrator	2017/10/17 16:19
8 C:\LabSolutions\Data\Dicamba 2-4 D\Optimization -connect.20171017_160752>Data.Dic...	-5	240	2	10	10	300	250	400	System Administrator	2017/10/17 16:20
9 C:\LabSolutions\Data\Dicamba 2-4 D\Optimization -connect.20171017_160752>Data.Dic...	-5	270	2	10	10	300	250	400	System Administrator	2017/10/17 16:22
10 C:\LabSolutions\Data\Dicamba 2-4 D\Optimization -connect.20171017_160752>Data.Dic...	-5	300	2	10	10	300	250	400	System Administrator	2017/10/17 16:23
11 C:\LabSolutions\Data\Dicamba 2-4 D\Optimization -connect.20171017_160752>Data.Dic...	-5	330	2	10	10	300	250	400	System Administrator	2017/10/17 16:25
12 C:\LabSolutions\Data\Dicamba 2-4 D\Optimization -connect.20171017_160752>Data.Dic...	-5	210	2	10	10	300	250	400	System Administrator	2017/10/17 16:27
13 C:\LabSolutions\Data\Dicamba 2-4 D\Optimization -connect.20171017_160752>Data.Dic...	-5	210	2.5	10	10	300	250	400	System Administrator	2017/10/17 16:28
14 C:\LabSolutions\Data\Dicamba 2-4 D\Optimization -connect.20171017_160752>Data.Dic...	-5	210	3	10	10	300	250	400	System Administrator	2017/10/17 16:30
15 C:\LabSolutions\Data\Dicamba 2-4 D\Optimization -connect.20171017_160752>Data.Dic...	-5	210	2.5	4	10	300	250	400	System Administrator	2017/10/17 16:32

Figure 5: LabSolutions Connect Source Optimization Results – Dicamba

Compound Name	Purity	Processor m/z	m/z1	CE1	Message	Interface Voltage	CID Gas	Nebulizing Gas	Drying Gas	Heating Gas	Interface Temp.	DL Temp	Heut Block Temp	User Name
1 2,4-D	-	219	161	13		-1	240	2.5	4	12	350	250	450	System Administrator



Data File Name	Interface Voltage	CID Gas	Nebulizing Gas	Drying Gas	Heating Gas	Interface Temp	DL Temp	Heut Block Temp	User Name	Date Acquired
1 C:\LabSolutions\Data\Dicamba 2-4 D\2,4-D\20171018_155458\Data\2,4-D_3vVoltage_1.icd	-1	270	2	10	10	300	250	400	System Administrator	2017/10/18 15:56
2 C:\LabSolutions\Data\Dicamba 2-4 D\2,4-D\20171018_155458\Data\2,4-D_3vVoltage_2.icd	-1.5	270	2	10	10	300	250	400	System Administrator	2017/10/18 15:57
3 C:\LabSolutions\Data\Dicamba 2-4 D\2,4-D\20171018_155458\Data\2,4-D_3vVoltage_3.icd	-2	270	2	10	10	300	250	400	System Administrator	2017/10/18 15:59
4 C:\LabSolutions\Data\Dicamba 2-4 D\2,4-D\20171018_155458\Data\2,4-D_3vVoltage_4.icd	-3	270	2	10	10	300	250	400	System Administrator	2017/10/18 16:00
5 C:\LabSolutions\Data\Dicamba 2-4 D\2,4-D\20171018_155458\Data\2,4-D_3vVoltage_5.icd	-4	270	2	10	10	300	250	400	System Administrator	2017/10/18 16:02
6 C:\LabSolutions\Data\Dicamba 2-4 D\2,4-D\20171018_155458\Data\2,4-D_3vVoltage_6.icd	-5	270	2	10	10	300	250	400	System Administrator	2017/10/18 16:03
7 C:\LabSolutions\Data\Dicamba 2-4 D\2,4-D\20171018_155458\Data\2,4-D_CIDGas_7.icd	-1	210	2	10	10	300	250	400	System Administrator	2017/10/18 16:05
8 C:\LabSolutions\Data\Dicamba 2-4 D\2,4-D\20171018_155458\Data\2,4-D_CIDGas_8.icd	-1	240	2	10	10	300	250	400	System Administrator	2017/10/18 16:07
9 C:\LabSolutions\Data\Dicamba 2-4 D\2,4-D\20171018_155458\Data\2,4-D_CIDGas_9.icd	-1	270	2	10	10	300	250	400	System Administrator	2017/10/18 16:08
10 C:\LabSolutions\Data\Dicamba 2-4 D\2,4-D\20171018_155458\Data\2,4-D_CIDGas_10.icd	-1	300	2	10	10	300	250	400	System Administrator	2017/10/18 16:10
11 C:\LabSolutions\Data\Dicamba 2-4 D\2,4-D\20171018_155458\Data\2,4-D_CIDGas_11.icd	-1	330	2	10	10	300	250	400	System Administrator	2017/10/18 16:11
12 C:\LabSolutions\Data\Dicamba 2-4 D\2,4-D\20171018_155458\Data\2,4-D_NebulizedGas_12.icd	-1	240	2	10	10	300	250	400	System Administrator	2017/10/18 16:13
13 C:\LabSolutions\Data\Dicamba 2-4 D\2,4-D\20171018_155458\Data\2,4-D_NebulizedGas_13.icd	-1	240	2.5	10	10	300	250	400	System Administrator	2017/10/18 16:15
14 C:\LabSolutions\Data\Dicamba 2-4 D\2,4-D\20171018_155458\Data\2,4-D_NebulizedGas_14.icd	-1	240	3	10	10	300	250	400	System Administrator	2017/10/18 16:16
15 C:\LabSolutions\Data\Dicamba 2-4 D\2,4-D\20171018_155458\Data\2,4-D_DryingGas_15.icd	-1	240	2.5	4	10	300	250	400	System Administrator	2017/10/18 16:18

Figure 6: LabSolutions Connect Source Optimization Results - 2,4-D




Figure 7: Limit of Detection for Dicamba (0.01 ppb), Injection 13




Figure 8: Limit of Detection for Dicamba (0.01 ppb), Injection 222




Figure 9: Limit of Detection for 2,4-D (0.001 ppb), Injection 13




Figure 10: Limit of Detection for 2,4-D (0.001 ppb), Injection 222




Figure 11: Limit of Quantitation for Dicamba (0.1 ppb), Injection 16




Figure 12: Limit of Quantitation for Dicamba (0.1 ppb), Injection 225




Figure 13: Limit of Quantitation for 2,4-D (0.01 ppb), Injection 16




Figure 14: Limit of Quantitation for 2,4-D (0.01 ppb), Injection 225

Table 1: % RSD of Spiked Formulation Injections

Standard/QC	Dicamba Conc. (ppb)	2,4-D Conc. (ppb)	Number of Injections	Dicamba % RSD	2,4-D % RSD
STD 1	10	1	2	6.72	8.77
STD 2	8	0.8	2	6.59	8.75
HQC	7.5	0.75	15	2.06	3.32
STD 3	5	0.5	2	6.22	6.40
MQC	5	0.5	134	1.87	3.16
STD 4	2.5	0.25	2	5.94	7.80
STD 5	0.5	0.05	2	7.41	10.92
LQC	0.3	0.03	15	3.13	5.78
STD 6	0.1	0.01	2	5.79	10.07
LOQ	0.1	0.01	15	3.23	5.73
LOD	0.01	0.001	15	10.5	5.80

Table 2: % Accuracy of both Dicamba and 2,4-D in Two Curves

Injection Number	Standard	Dicamba Area	Dicamba % Accuracy	2,4-D Area	2,4-D % Accuracy
5	STD 6	157720	105.68	155675	106.57
6	STD 5	848221	91.81	750810	119.08
7	STD 4	4822282	99.71	3248143	105.87
8	STD 3	10205788	104.96	6165049	100.85
9	STD 2	17256649	110.70	10058179	103.01
10	STD 1	22038727	113.04	12691910	104.04
178	STD 6	145319	99.34	134981	89.57
179	STD 5	763722	83.16	643179	101.40
180	STD 4	4433748	91.76	2908383	94.70
181	STD 3	9346019	96.16	5631124	92.08
182	STD 2	15720144	100.87	8886042	90.97
183	STD 1	20038771	102.80	11209539	91.86




Figure 15: Dicamba Curve and QCs




Figure 16: 2,4-D Curve and QCs




Figure 17: Dicamba Mid QC Areas, 134 Injections




Figure 18: 2,4-D Mid QC Areas, 134 Injections




Figure 19: Dicamba LOD Areas, Injection 4 through 222




Figure 20: 2,4-D LOD Areas, Injection 4 through 222

■ Results and Discussion

The Shimadzu LCMS-8060 showed sufficient S/N at ultralow detection limits of 0.01 ppb and 0.0001 ppb for dicamba and 2,4-D, respectively. The area counts of four replicate injections of 0.1 ppb dicamba had a %RSD of 4.04.

The Shimadzu LCMS-8060 showed good repeatability and sensitivity over approximately 24 hours of continuous injections. The % RSD was below 11% across all levels including the LOD. The mid-QC areas range from 9.3 to 10.0 million for dicamba and 5.5 to 6.2 million for 2,4-D. These areas are from injection 22 to 230 (5 minute run time). Even the LOD injections have great % RSDs from 10.5 to 5.8% for dicamba and 2,4-D, respectively.

Even with a 100x dilution, it is apparent that the samples are still quite dirty upon injection. The robust interface design of the LCMS-8060 with its removable Desolvation Line will allow analysts to perform routine source cleanings, without breaking vacuum, in order to maintain the instrument's peak operating condition for this ultra-sensitive analysis.

■ Conclusion

The LCMS-8060 has proven to be a very sensitive and robust platform for this application. The method as it stands appears sufficient for routine herbicide analysis.

UPLC-MS

ULTRA FAST MASS SPECTROMETRY



LCMS-8040



LCMS-8045



LCMS-8050



LCMS-8060



LCMS-2020



LCMS-IT-TOF

Founded in 1875, Shimadzu Corporation, a leader in the development of advanced technologies, has a distinguished history of innovation built on the foundation of contributing to society through science and technology. Established in 1975, Shimadzu Scientific Instruments (SSI), the American subsidiary of Shimadzu Corporation, provides a comprehensive range of analytical solutions to laboratories throughout North, Central, and parts of South America. SSI maintains a network of nine regional offices strategically located across the United States, with experienced technical specialists, service and sales engineers situated throughout the country, as well as applications laboratories on both coasts.

For information about Shimadzu Scientific Instruments and to contact your local office, please visit our Web site at www.ssi.shimadzu.com



Shimadzu Corporation
www.shimadzu.com/an/

SHIMADZU SCIENTIFIC INSTRUMENTS, INC.
Applications Laboratory
7102 Riverwood Drive, Columbia, MD 21045
Phone: 800-477-1227 Fax: 410-381-1222
URL <http://www.ssi.shimadzu.com>

For Research Use Only. Not for use in diagnostic procedures. The content of this publication shall not be reproduced, altered or sold for any commercial purpose without the written approval of Shimadzu. The information contained herein is provided to you "as is" without warranty of any kind including without limitation warranties as to its accuracy or completeness. Shimadzu does not assume any responsibility or liability for any damage, whether direct or indirect, relating to the use of this publication. This publication is based upon the information available to Shimadzu on or before the date of publication, and subject to change without notice.

© Shimadzu Scientific Instruments, 2017
First Edition: December 2017